

CORRIGENDUM: e-Tender ID No: 2025_ARDD_63151_1, published on: 26/06/2025



GOVERNMENT OF TRIPURA
OFFICE OF THE PRINCIPAL
COLLEGE OF VETERINARY SCIENCES & ANIMAL HUSBANDRY R.K.
NAGAR, WEST TRIPURA

F.No.4-46/CVS/STORE/INST/RKN/23-24

Dated: _____, 2025

CORRIGENDUM

This is hereby informed that the following corrections has been made in the tender document vide tender ID No: 2025_ARDD_63151_1, published on: 26/06/2025 in connection with the **e-Tender (2nd Call)** for procurement of remaining **35 (thirty five)** numbers of laboratory instruments, equipment, appliances for the College of Veterinary Sciences and Animal Husbandry under the NEC funded project for the Financial Year: 2025 – 2026

Sl. No	Page Number of D-eNIT	Particulars of corrigendum
1	Page-21 at Sl. No 3 of Annexure-III	The specification of BOD incubator appear at Serial Number: 3 of published De-NIT has been corrected as follows: <ul style="list-style-type: none"> - Temperature Range: Ambient +5°C to 60°C - Temperature Accuracy : ±0.5°C - Temperature sensor: PT-100, with compressor. - Capacity: 200 Liters. - No. of Trays: 3 - No. of Doors: Single - Lighting: Interior illumination with LED Light. - Power: Works on 230 VAC single phase 50Hz. - Warranty:-3 Years (Onsite)
2	Page Number -22 at Sl. No. 4 Of Annexure-III	The video Endoscopy machine must be supplied with 1-KVA-UPS.
3.	Page Number -23 at Sl. No. 6 Of Annexure-III	Bacteriological Incubator temperature range may be read as Ambient +5°C to 60°C
	The Last date & time for submission of e-bids has been extended till 04:00 pm of 24.07.2025. and the date & time for opening of e-tender will be on 25.07.2025, time: office hours.	

The other specifications and terms & conditions of D-eNIT will remain same.

Signed by Purabi Barman

Date: 15-07-2025 20:11:57

Principal,

College of Veterinary Sciences &
Animal Husbandry,
R.K. Nagar, West Tripura